

VPDUSS-SOP-8.5.1.2-01.F02

VAISHAL PATLIPUTRA DUGDH UTPADAK SAHKARI SANGH LTD
PATNA DAIRY PROJECT
PHULWARISHARIF, PATNA 801 505
PHONE :0612-2252542, 2252553, 2251622, E-mail : vpmu.pur@gmail.com

PDP: PUR: 7683

07.03.2024

8

Dear sir,

We are interested in purchasing of the following items. In case you can supply the same, you may send your quotation in sealed cover addressed to the Managing Director address given above superscribing on the top of the envelope quotation for vet. Lab items so as to reach us on or before 24.03.2024.

All rates must be F.O.R.- Hajipur on door delivery basis.

S.No.	Item	Qty reqd
1	Bio-chemical Analyser As per specification enclosed no 3	1 nos

Thanking you,

Yours faithfully,


Manager (MM)

Annexure-I

Technical Specification for semi-automated biochemical analyzer

- 20°C – 40°C range of temperature selection
- Photometric range: 350-670 or better
- Interference filters: 7 or more
- Triple cuvette system or better
 - 20 µl or less Flow cell
 - 12 mm or less Square cuvette
- 7 mm or less round glass tube with adaptor
- Peltier temperature for 25°C-40°C or better
- Minimum aspiration volume 220 µl or less

Memory:

- 180 or more User defined test programs
- 900 or more test result
- LCD display with High resolution, graphics
- In built high resolution graphic thermal printer

(3)